

U.S. Bid Specifications

Automated External Defibrillator (AED)

HeartSine® samaritan® PAD SAM 350P or SAM 360P

HeartSine Connected AED SAM 350P or SAM 360P with LIFELINKcentral™ AED program manager

This document provides features of the HeartSine samaritan PAD 350P and 360P. Specifications are the same for both models, except where noted. Additional features of connected versions of the AEDs are identified separately.

Defibrillator	
In package	<ul style="list-style-type: none"> HeartSine samaritan PAD is supplied with an adult Pad-Pak (combination electrodes and battery), operating instructions and carry case, as standard. Each HeartSine Connected AED also includes a HeartSine Gateway, multilingual HeartSine Gateway Set-up guide, batteries, removal tool and LIFELINKcentral AED program manager, Basic license. NOTE: The complete user manual for HeartSine Gateway is available online at heartsine.com.
Waveform	<ul style="list-style-type: none"> Optimized, impedance-based, biphasic truncated exponential, escalating waveform, called Self-Compensating Output Pulse Envelope (SCOPE), which compensates voltage, slope (tilt) and duration for patient impedance.
Patient impedance range	<ul style="list-style-type: none"> 20 – 230 ohms.
Patient analysis system	<ul style="list-style-type: none"> Evaluates patient’s ECG, electrode contact integrity and patient impedance to determine if defibrillation is required.
Energy protocol	<ul style="list-style-type: none"> Adult: 150J, 150J, 200J Pediatric: 50J, 50J, 50J
Time to shock following CPR (150J)	<ul style="list-style-type: none"> SAM 350P: Typically, 8 seconds SAM 360P: Typically, 19 seconds
Charge time	<ul style="list-style-type: none"> Typically, 150J in < 8 seconds, 200 J in < 12 seconds
Warranty	<ul style="list-style-type: none"> HeartSine AED is backed by an 8-year limited warranty.
Guideline recommendations	<ul style="list-style-type: none"> HeartSine AED follows the recommendations in the latest AED guidelines for the American Heart Association (AHA), European Resuscitation Council (ERC) and International Liaison Committee on Resuscitation (ILCOR). HeartSine AED is consistent with current guidelines for CPR (energy protocols, CPR protocols, quality assurance).

Operation

Semi-auto vs. Fully auto	<ul style="list-style-type: none"> • HeartSine SAM 350P: Upon delivery, the HeartSine samaritan PAD 350P is configured for semi-automatic operation (shock button press needed to deliver a shock). • HeartSine SAM 360P: Upon delivery, the HeartSine samaritan PAD 360P is configured for fully automatic operation (shock is delivered without a button press).
On/Off	<ul style="list-style-type: none"> • HeartSine AED powers on when the ON/OFF button is pressed.
Manual override	<ul style="list-style-type: none"> • Manual override is not available on the HeartSine AED.
Intended use	<ul style="list-style-type: none"> • The devices are intended for use by personnel who have been trained in their operation. Training on CPR and in the use of an AED is strongly recommended for users. However, in an emergency situation the HeartSine samaritan PAD may be used by an untrained lay rescuer.
Voice prompts	<ul style="list-style-type: none"> • HeartSine AED guides the operator through operating procedures with a combination of voice prompts, flashing LEDs and visual prompts.
Metronome	<ul style="list-style-type: none"> • HeartSine AED provides an audible metronome that sounds at a rate of between 100 to 120 beats per minute, as per the current guidelines.
Electrode placement	<ul style="list-style-type: none"> • HeartSine AED provides visual prompts (pictures) for placement of the electrodes: <ul style="list-style-type: none"> – Anterior-lateral (Adult) – Anterior-posterior or Anterior-lateral (Pediatric).
CPR prompts	<ul style="list-style-type: none"> • HeartSine AED prompts the operator to perform CPR for 120 seconds.
CPR feedback	<ul style="list-style-type: none"> • HeartSine SAM 350P and SAM 360P provide CPR coaching, but do not provide CPR feedback.
Child mode	<ul style="list-style-type: none"> • Upon power on when an adult Pad-Pak is inserted, the HeartSine AED defaults to the energy delivery and CPR guidance set for an adult patient. • When a Pediatric-Pak is inserted (either before or after the device is powered on), the HeartSine AED defaults to the energy delivery and CPR guidance set for a pediatric patient.
Electrodes	<ul style="list-style-type: none"> • The HeartSine AED uses an interchangeable battery and electrode pack called Pad-Pak. The HeartSine AED in combination with an adult Pad-Pak is suitable for use on patients of over 55 lb (25 kg) in weight or equivalent to a child of approximately eight years old or over. • For use on smaller children (from 1 to 8 years old), a Pediatric-Pak is used. If a Pediatric-Pak or an alternative suitable defibrillator is not available, an adult Pad-Pak may be used.

Real-time clock	<ul style="list-style-type: none"> HeartSine AED has a real-time clock that uses Coordinated Universal Time (UTC).
Readiness indicator	<ul style="list-style-type: none"> The Readiness indicator is a bright green LED that flashes every five to ten seconds to indicate the HeartSine AED is ready for use. If the HeartSine AED has detected an issue affecting readiness, the readiness indicator light on the HeartSine AED will flash red (approximately every 5-10 seconds) and emit an audible beep. For the HeartSine Connected AED, if the HeartSine AED has detected an issue affecting readiness, after the device performs the weekly check-in, the Not Ready condition also will be reported to LIFELINKcentral AED program manager via a Wi-Fi connection and an email is sent to the AED program manager, if set up in LIFELINKcentral.
Low battery indication	<ul style="list-style-type: none"> When the battery needs to be replaced, the indicator light on the HeartSine AED will flash red (approximately every 5-10 seconds) and emit an audible beep.
Accidental switch-off	<ul style="list-style-type: none"> If the HeartSine AED is switched off during shock advisory, a warning prompt is heard and a confirmation button press is required to proceed to turn off the AED.
Self-test (self-check)	<ul style="list-style-type: none"> HeartSine AED performs an automatic self-test on a weekly basis. If an automatic self-test detects a condition that requires attention, this negative result of the self-test is indicated by visible and audible indicators on the device. HeartSine AED also runs a self-test each time the device is turned on. For the HeartSine Connected AED, this negative result is reported in LIFELINKcentral AED program manager with an alert sent to the pre-defined AED program administrator(s). If automatic self-tests indicate the device is ready, HeartSine Connected AED checks in to LIFELINKcentral AED Program Manager once each month and reports that it is READY.
Use on aircraft	<ul style="list-style-type: none"> HeartSine AED is suitable for use and storage on airplanes when used in conjunction with the TSO/ETSO-approved battery and electrodes (Pad-Pak-07), made specifically for use on aircraft. The manufacture and use of SAM 350/350P models comply with these aviation standards for use on aircraft: <ul style="list-style-type: none"> RTCA / DO-160G Environmental Conditions and Test Procedures for Airborne Equipment, Section 21: Emission of Radio Frequency Energy (Category M)

Physical Specifications	
Handle/case	<ul style="list-style-type: none"> • Each HeartSine AED is shipped with a carrying case.
Dimensions	<ul style="list-style-type: none"> • HeartSine AED with Pad-Pak: 8.0 in x 7.25 in x 1.9 in (20 cm x 18.4 cm x 4.8 cm) • HeartSine Connected AED with Pad-Pak, HeartSine Gateway and batteries: 9.21 in x 7.25 in x 1.9 in (23.4 cm x 18.4 cm x 4.8 cm)
Weight	<ul style="list-style-type: none"> • HeartSine AED with Pad-Pak: 2.4 lb (1.1 kg) • HeartSine Connected AED with Pad-Pak, HeartSine Gateway and batteries: 2.83 lb (1.285 kg)
Display	<ul style="list-style-type: none"> • Easy-to-understand visual and voice prompts guide the rescuer through the entire resuscitation process, including CPR
Port	<ul style="list-style-type: none"> • HeartSine AED: Proprietary USB data port connects to HeartSine USB data cable • HeartSine Connected AED: Micro USB data port
Communications	<ul style="list-style-type: none"> • HeartSine AED: Connection to Saver EVO software through a data port that uses a custom USB data cable. • HeartSine Connected AED: Communication is via Wi-Fi 802.11 b/g/n data transfer to LIFELINKcentral AED program manager or through a USB connection to Saver EVO software through a micro USB port.
Patient Analysis System	
ECG analysis	<ul style="list-style-type: none"> • The HeartSine AED evaluates patient's ECG, electrode contact integrity and patient impedance to determine if defibrillation is required. • The analysis is performed on both adult and pediatric patients.
Motion detection	<ul style="list-style-type: none"> • HeartSine samaritan PAD 360P fully automatic AED provides motion detection.
Overall sensitivity	<ul style="list-style-type: none"> • For shockable VF in excess of 95% • For shockable VT in excess of 90%.
Overall specificity	<ul style="list-style-type: none"> • For all non-shockable rhythms in excess of 95%.
Patient impedance	<ul style="list-style-type: none"> • HeartSine AED can deliver a therapeutic shock over a patient impedance range of 20 ohms to 230 ohms, without a significant reduction in delivered shock energy.

Environmental	
Operating temperature range	<ul style="list-style-type: none"> • 32°F to 122°F (0°C to 50°C)
Transport temperature range	<ul style="list-style-type: none"> • 32°F to 122°F (0°C to 50°C) <p>NOTE: It is recommended that the device should be placed in an ambient temperature of between 0°C to 50°C (32°F to 122°F) for at least 24 hours upon first receipt.</p>
Long term storage temperature range	<ul style="list-style-type: none"> • 32° to 122°F (0°C to 50°C)
Relative humidity	<ul style="list-style-type: none"> • 5 to 95% (non-condensing)
Atmospheric pressure (altitude)	<ul style="list-style-type: none"> • 0 to 15,000 feet (0 to 4,575 meters)
Water resistance	<ul style="list-style-type: none"> • Meets IEC 60529/EN 60529 IPX6 with electrodes connected and battery installed.
Dust resistance	<ul style="list-style-type: none"> • Meets IEC 60529/EN 60529 IP5X with electrodes connected, and battery installed.
Shock	<ul style="list-style-type: none"> • Meets MIL-STD-810F Method 516.5, Procedure 1 (40 G's).
Vibration	<ul style="list-style-type: none"> • Meets MIL-STD-810F Method 514.5, Procedure 1 • Category 4 Truck Transportation – US Highways • Category 7 Aircraft – Jet 737 & General Aviation
Drop test	<ul style="list-style-type: none"> • 3.3 feet (1 meter)
Event documentation	
Memory type	<ul style="list-style-type: none"> • Internal digital memory (flash RAM).
ECG storage	<ul style="list-style-type: none"> • Each HeartSine AED can store 90 minutes of ECG (full disclosure) and event/incident recording.
Data retrieval	<ul style="list-style-type: none"> • Event data is transferred using a custom USB data cable (non-connected devices) or a micro USB cable (connected devices) to a PC with Saver EVO Windows-based data review software.
Report software	<ul style="list-style-type: none"> • Saver EVO Windows-based data review software.

Batteries	
Type	<ul style="list-style-type: none"> HeartSine AED uses non-rechargeable, single-use combined defibrillation electrode and battery cartridge (Pad-Pak or Pediatric-Pak). The battery is lithium manganese dioxide (Li/MnO₂) 18V. HeartSine Gateway (on the HeartSine Connected AED) uses four CR123A 3V, non-rechargeable batteries.
Amp-hours – AED battery	<ul style="list-style-type: none"> 1.5
Capacity – New AED battery	<ul style="list-style-type: none"> Fully charged battery typically provides > 60 shocks at 200J or 6 hours of battery use.
Standby life (assuming only weekly tests)	<ul style="list-style-type: none"> HeartSine AED Battery: 4 years from manufacture date. (See the expiration date on the Pad-Pak/Pediatric-Pak.)
Weight	<ul style="list-style-type: none"> HeartSine AED Battery: 0.44 lb (0.2 kg) HeartSine Gateway Battery: 17 g (per battery)
Replacement cycle	<ul style="list-style-type: none"> All batteries should be replaced every four years if the device has not been used. Upon immediate expiration of the HeartSine AED battery contained in the Pad-Pak, there are at least 10 shocks at 200J (under normal device use).
Electrodes	
Adult / pediatric	<ul style="list-style-type: none"> HeartSine AED uses an interchangeable battery and electrode cartridge called Pad-Pak. The HeartSine AED in combination with an adult Pad-Pak is suitable for use on patients of over 55 lb (25 kg) in weight or equivalent to a child of approximately eight years old or over. For use on smaller children (from 1 to 8 years old), a Pediatric-Pak is used. If a Pediatric-Pak or an alternative suitable defibrillator is not available, an adult Pad-Pak may be used.
Operation	<ul style="list-style-type: none"> Electrodes are stored in a sealed protective package within the Pad-Pak or Pediatric Pak. The user pulls the green handle to reveal the electrodes package. The user simply opens the package to reveal the electrodes. The user then easily peels the electrodes from the tray and follows the image on each which indicates where it should be placed on the patient's chest.
Replacement cycle	<ul style="list-style-type: none"> Electrodes (Pad-Pak, Pediatric-Pak) must be replaced every 4 years if not used.
Electrode cable length	<ul style="list-style-type: none"> 3.3 feet (1 meter)

Accessories

<p>Defibrillator accessories</p>	<ul style="list-style-type: none"> • Several accessories are available, including the following: <ul style="list-style-type: none"> – Pad-Pak battery and electrode cartridge – Pad-Pak battery and electrode cartridge with ETSO/TSO-C142a certification for use on aircraft – Pediatric-Pak battery and electrode cartridge – HeartSine Gateway – Custom USB data cable – Mobile AED rescue backpack – Wall bracket – Rotaid Plus wall cabinet with alarm – Rotaid Solid Plus wall cabinet with alarm – Rotaid Solid Plus Heat wall cabinet with heat and alarm – Wall cabinet with alarm (rectangular) – HeartSine samaritan PAD carry case (replacement)
<p>Trainer/ Trainer accessories</p>	<ul style="list-style-type: none"> • HeartSine samaritan PAD Trainer (available for each HeartSine AED model) • Trainer-Pak (replacement electrode cartridge) • Replacement Trainer electrodes (for use with Trainer-Pak) • Remote (replacement) • Battery charger (replacement)
<p>On-line training</p>	<ul style="list-style-type: none"> • On-line AED training is available to provide instructions on operating the device.

Trainer

Description	<ul style="list-style-type: none"> HeartSine samaritan PAD Trainer provides realistic training in the use of the HeartSine AED without the actual charge and discharge of electrical energy. Designed to simulate the appearance of the HeartSine AED, HeartSine Trainer simulates its actual operation using the same simple user interface. HeartSine Trainer is compatible with any CPR manikin system.
Labeling	<ul style="list-style-type: none"> Includes clear labeling to identify it as a training unit, including a gold face.
Dimensions	<ul style="list-style-type: none"> 8.0 in x 7.25 in x 1.9 in (20 cm x 18.4 cm x 4.8 cm)
Weight (with batteries)	<ul style="list-style-type: none"> 1.3 lb (0.6 kg)
Buttons	<ul style="list-style-type: none"> On/Off, Shock
Simulation	<ul style="list-style-type: none"> All of the functions of HeartSine Trainer are programmable to simulate the full range of scenarios users might encounter during “real life” use of the HeartSine AED. HeartSine Trainer provides six pre-set AHA/ERC scenarios.
Training electrodes	<ul style="list-style-type: none"> The training electrodes are reusable and are quick and easy to replace back into the electrode tray.
Battery	<ul style="list-style-type: none"> HeartSine Trainer uses a rechargeable battery which will provide approximately seven hours of usage before needing to be recharged. A low battery message will be played approximately one hour prior to the battery becoming depleted.

AED program management

<p>LIFELINKcentral AED program manager</p>	<ul style="list-style-type: none"> • LIFELINKcentral AED program manager enables managing readiness of a HeartSine Connected AED and HeartSine Gateway via Wi-Fi connectivity to the attached HeartSine Gateway. The program also enables manual management of AED location and Pad-Pak and Pediatric-Pak expirations, as well as of non-connected AEDs.
<p>Licenses and service packages</p>	<ul style="list-style-type: none"> • Basic license, included at no extra charge with every HeartSine Connected AED and HeartSine Gateway, enables remote management to monitor equipment and site readiness and to manually track expirations of Pad-Pak and Pediatric-Pak, providing immediate alerts if the device needs attention. Email support and access to the online Resource Center are also provided. • The PRO license, available per account for a nominal fee with a HeartSine AED (connected or non-connected) and HeartSine Gateway, offers full portal access with expanded view and functionality: <ul style="list-style-type: none"> – Remote monitoring of equipment and site readiness – Manual tracking of non-connected Stryker AEDs – Manual tracking of Pad-Pak and Pediatric-Pak expirations – Immediate alerts if the device needs attention – Manual tracking/readiness notifications for non-Stryker devices, including other AEDs, accessories and disposables, and other safety devices – Manual management of training events and rosters, plus manual tracking of responder readiness (CPR/AED certifications) – Viewing of equipment, site and training readiness wheels – Setting site customer inspection schedules and recording of manual inspections <ul style="list-style-type: none"> – Email support and access to the online Resource Center • Service packages are available in some countries. See the LIFELINKcentral AED program manager brochure for details on services provided.

BRIEF SUMMARY OF INDICATIONS AND IMPORTANT SAFETY INFORMATION

INDICATIONS FOR USE: The HeartSine samaritan PAD SAM 350P (SAM 350P), HeartSine samaritan PAD SAM 360P (SAM 360P) and HeartSine samaritan PAD SAM 450P (SAM 450P) are indicated for use on victims of cardiac arrest who are exhibiting the following signs: unconscious, not breathing, without circulation (without a pulse). The devices are intended for use by personnel who have been trained in their operation. Users should have received training in basic life support/AED, advanced life support or a physician-authorized emergency medical response training program. The devices are indicated for use on patients greater than 8 years old or over 55 lb (25 kg) when used with the adult Pad-Pak™ (Pad-Pak-01 or Pad-Pak-07). They are indicated for use on children between 1 and 8 years of age or up to 55 lb (25 kg) when used with the Pediatric-Pak™ (Pad-Pak-02).

CONTRAINDICATION: If the patient is responsive or conscious, do not use the HeartSine samaritan PAD to provide treatment.

WARNINGS: AEDs: • The HeartSine samaritan PAD delivers therapeutic electrical shocks that can cause serious harm to either users or bystanders. Take care to ensure that no one touches the patient when a shock is to be delivered. • Touching the patient during the analysis phase of treatment can cause interference with the diagnostic process. Avoid contact with the patient while the HeartSine samaritan PAD is analyzing the patient. The device will instruct you when it is safe to touch the patient. • Do not delay treatment trying to find out the patient's exact age and weight. If a Pediatric-Pak or an alternative suitable defibrillator is not available, you may use an adult Pad-Pak. • The SAM 360P is a fully automatic defibrillator. When required, it will deliver a shock to the patient WITHOUT user intervention. • The SAM 450P CPR Rate Advisor is currently only intended to provide feedback on adult patients. If you treat a pediatric patient with the SAM 450P and an adult Pad-Pak, ignore any voice prompts regarding the rate of CPR. • Do NOT use the HeartSine samaritan PAD in the vicinity of explosive gases, including flammable anesthetics or concentrated oxygen. • Do NOT open or repair the device under any circumstances as there could be danger of electric shock. If damage is suspected, immediately replace the HeartSine samaritan PAD. **Pad-Paks:** • Do not use if the gel is dry. • The Pediatric Pad-Pak is not for use on patients under 1 year old. For use with children up to the age of 8 years or up to 55 lb (25 kg). DO NOT DELAY THERAPY IF YOU ARE NOT SURE OF EXACT AGE OR WEIGHT. • Only HeartSine samaritan PADs with the label are suitable for use with the Pediatric-Pak. If the HeartSine samaritan PAD you are using does not have this label, use the adult Pad-Pak if no alternatives are available. • The use of the Pediatric-Pak will enable delivery of 50J shocks to the pediatric patient. • The Pediatric-Pak contains a magnetic component (surface strength 6500 gauss). Avoid storage next to magnetically sensitive storage media. It is advised that Pediatric-Paks are stored separately when not in use. • Never charge, short circuit, puncture, deform, incinerate, heat above 85oC or expose contents of TSO (Aviation) Pad-Pak to water. Remove when discharged.

PRECAUTIONS: AEDs: • Proper placement of the HeartSine samaritan PAD electrode pads is critical. Electrode pads must be at least 1 in (2.5 cm) apart and should never touch one another. • Do not use electrode pads if pouch is not sealed. • Check the device periodically in accordance with the service and maintenance instructions provided in the User Manual. • Operate the HeartSine samaritan PAD at least 6 feet (2 meters) away from all radio frequency devices or switch off any equipment causing interference. • Use of the device outside the operating and storage ranges specified in the User Manual may cause the device to malfunction or reduce the shelf life of the Pad-Pak. • Do not immerse any part of the HeartSine samaritan PAD in water or any type of fluid. • Do not turn on the device unnecessarily as this may reduce the standby life of the device. • Do not use any unauthorized accessories with the device as the HeartSine samaritan PAD may malfunction if non-approved accessories are used. • Dispose of the device in accordance with national or local regulations. • Check with the relevant local government health department for information about any requirements associated with ownership and use of a defibrillator in the region where it is to be used. **Pad-Paks:** • Check expiration date. **Saver EVO™ Software:** • Download the complete HeartSine samaritan PAD memory prior to erasing it. This information should be stored safely for future reference. Ensure that only the events you want to delete have been selected prior to deleting. Once deleted from your computer's memory, events cannot be regenerated and all information will be lost.

POTENTIAL ADVERSE EFFECTS: The potential adverse effects (e.g., complications) associated with the use of an automated external defibrillator include, but are not limited to, the following: • Failure to identify shockable arrhythmia. • Failure to deliver a defibrillation shock in the presence of VF or pulseless VT, which may result in death or permanent injury. • Inappropriate energy which could cause failed defibrillation or post-shock dysfunction. • Myocardial damage. • Fire hazard in the presence of high oxygen concentration or flammable anesthetic agents. • Incorrectly shocking a pulse-sustaining rhythm and inducing VF or cardiac arrest. • Bystander shock from patient contact during defibrillation shock. • Interaction with pacemakers. • Skin burns around the electrode placement area. • Allergic dermatitis due to sensitivity to materials used in electrode construction. • Minor skin rash.

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a physician.

Please consult the User Manual at www.heartsine.com for the complete list of indications, contraindications, warnings, precautions, potential adverse events, safety and effectiveness data, instructions for use and other important information.

©2021 Stryker. All names herein are trademarks or registered trademarks of their respective owners. M0000002297-AB

FOR INTERNAL/DISTRIBUTOR USE ONLY